

Rule Number	Rule change Requested	Requester	Status	Revision
333-111-0015	Revise OAR 333-111-0015 to meet 10 CFR Part 19, Annual reports to workers if they exceed 100 mrem TEDE or 100 mrem	Rainier Farmer	Change to meet RATS ID 2008-1	2010-2
333-101	Minimum licensing requirements and fee for health physics consultant	Carpenter	Review for addition	2010-2
333-101-0023(1)	Rule needs to capture Radon Service Vendor	Terry	Change for next revision	2010-2
333-101-0035(3)(4)	Delete (3) and (4).	Smith	Review for change - in statute	2011
333-101-0065	Change "person" to "machine"		Housekeeping	2010-2
333-101-0090	Investigate the X-ray civil penalty rule. Is it compatible to Division 124?	Todd	Review for change	2010-2
Division 102	Change radioactive drugs to radiopharmaceutical drugs and add the term Byproduct Material	Todd	Correction made to draft	
333-102-0115(9)	Numbering, house keeping	Siebert	Correction made to draft	2010-2
333-102-0190(8)	Add "or" in 333-102-0190 and add "the device; and" in (8)(c)(A)	RATS 2007-3	Correction made to draft	2010-2
333-102-0285	See comment #4 of comment letter	RATS 2007-3	Correction made to draft	2010-2
333-102-0290(1)	Added one rule (333-116-0485 to match 35.1000) to 333-102-0290(1)	RATS 2007-1	Correction made to draft	2010-2
333-102-0305(23)	Decay in storage to match CFR 35.92 and RATS 2007-1	RATS 2007-1	Correction made to draft	2010-2
333-102-0305(28)	Re word 3 security controls	Justin	Review for change	2010-2
333-102-0305(30)	333-102-0305(30) is not equivalent to 10 CFR 32.72.	RATS 2007-3	Correction made to draft	2010-2
333-102-0305(30)(a) & (b)	Change cross reference rules	RATS 2007-3	Correction made to draft	2010-2
333-102-0340	Investigate if RML should remove (6). "Shall notify the department within one hour after arrival at the actual work site.	Kevin	Review for change	2010-2
333-102-0340	Remove 180 days for reciprocity	Carpenter	Review for change	2011
333-102-0340(1)(b)	Grounds to deny reciprocity	Spence	Text added to rule	2010-2
333-102-0340(2)	Add rule to section "Pay applicable fees for filing reciprocity registration per OAR 333-103 0010(2)(b) for vendors and services licensees or appropriate fee for license type listed in OAR 333-103-0010	Lindsey	Review for change	2010-2
333-103-0010(2)(t)	Add the statement "For medical licensees with more than 7 sources, may apply for a basic license for attenuation purposes.	Lindsey and Peter Chin	Text added to rule	2010-2
333-103-0035(2)©	Need to change "c" with "C"	Siebert	House keeping	2010-2
333-105-0530(1)(a)	Consider the deletion of "Oregon Regulatory Guide, ORG-2, Industrial Radiography".	Carpenter	Correction made to draft	2010-2
333-106-0035	Add estimated patient dose requirement to variance request matrix	Terry	Change for next revision	2010-2
333-106-0045(3)	Explore the definition of "impractical". Perhaps replace with the meaning of "Limited Access Patient" or add this as a new	ODA	ODA	2011

333-106-0055	Rule reference to the Oregon Board of Radiologic Technologist will need to change to "Oregon Board of Medical Imaging". Need to check OARs and ORS for name change throughout the division.	Margaret Lut	Correction made to draft	2010-2
333-106-0055(1)	Add (1)(k) after (1)(j) to the statement "not required for Greens ray X-ray equipment operator training. Also, correct an to a Department approved course.	Crawford	Correction made to draft	2010-2
333-116	Change Certification to Attestation and radioactive drug to radiopharmaceutical drug	Carpenter	Housekeeping	2010-2
333-116-0035	Need to examine rule language for HDR. Confusion with statement "Except for medical use of remote afterloaders, a separate license application must be filed for each medical use of radioactive materials as described in OAR 333-116-0480". Makes it sound that HDR does not require a separate	Sylvia and Terry	Text revised to communicate that a separate license must be submitted for HDRs.	2010-2
333-116-0105	Add I-125 to rule	Spence	Text added to rule	2010-2
333-116-0250(3)	Conversion from traditional units to SI units is incorrect	Justin	Correction made to draft	2010-2
333-116-0260(1)	Revised to meet 35.75 Added note to meet CFR	RATS 2007-1	Correction made	2010-2
333-116-0290	10 CFR 35 vs. 333-116-0290 & 333-102-0305(23)	Martin	Review for addition	2010-2
333-116-0290	Decay in storage to match CFR 35.92 and RATS 2007-1	RATS 2007-1	Correction made	2010-2
333-116-0330	333-116-0330. Oregon did not revise to include the equivalent for strontium-82 found in 10 CFR 35.204	RATS 2007-3	Correction made to draft	2010-2
333-116-0390	Symbols incorrect	Justin	Correction made to draft	2010-2
333-116-0470	Conversion from traditional units to SI units	Justin	Correction made to draft	2010-2
333-116-0660(3)	Added "include subsections (a) through (b):"	RATS	Correction made	2010-2
333-116-0670	Delete 333-116-0670(a) and (b). Change 116-0670(b) to 116-0670(a).	RATS 2007-1	Rule housekeeping	2010-2
333-116-0670(3)	Revised wording	RATS 2007-1	Correction made to draft	2010-2
333-119-0020(16)	Define "Open public area" Create new definition 119-0020(16)	RAC	Correction made to draft	2010-2
333-119-0080(3) & (4)	Reword "On or after July 1, 2011, all operators of registered tanning	Shannon	Change for next revision	2011
333-119-0080(3)(a) and (b)	Delete rules regarding beds vs. training operator requirement	Molly	Change for next revision	2011
333-119-0100(20)	Remove the term " Commencing January 1, 2011 " Rule will read; Tanning devices intended for commercial use that were manufactured prior to September 8, 1986 shall be prohibited	Terry	Correction made to draft	2010-2

333-120-0545	Contact DOE and verify that the language is correct	RACE	Review for change	2010-2
333-120-0550(5)	Need to insert the phrase "in accordance with 10 CFR Part 20, Appendix G" after ...to the intended consignee	RATS 2007-3	Correction made to draft	2010-2
333-123-0040(2)	Emergency shutoff switch.	Wright	Review for change	2011
Create section for CT operations			Lut	2011
Division 102 and 116	Change radioactive drugs to radiopharmaceutical drugs	RAC	Correction made	2010-2
Division 106	New rule to define permanent clinic vs temporary	ODA	ODA	2011
RML Divisions	Revise rules concerning Radioactive Material Licensing security updates for background, fingerprinting, TR appointment and SUNSI.	Todd	Change next revision	2010-2
Unknown division	Allow veterinary to use hand-held X-ray unit		Review for change	2010-2
Various	Need to add total "dose" for variance request.	Terry	Change for next revision	2010-2